

CERTIFICATE OF ANALYSIS

4200 Research Forest Drive, Suite 150
The Woodlands, TX 77381
Phone: (281) 298-9410 Fax: (281) 298-9411

FINISHED PRODUCT: PolySeed NX[®]

LOT NUMBER: 160603-1

FORMULATION:

The formulation for this product contains a range of naturally occurring microorganisms, which are known to be non-pathogenic to man or animals.

VIABLE COUNT, FINAL TEST RESULT:

The product has been fully tested in accordance with Finished Product Specifications and contains a minimum viable count of 4.00×10^9 cfu/g

GLUCOSE/GLUTAMIC-ACID RESULTS:

Tested results within acceptable range 198 +/- 30.5 mg/L (167.5 - 228.5 mg/L).
See www.polyseed.com for details.

SEED CONTROL FACTOR:

Tested results within acceptable range 0.60 - 01.0. See www.polyseed.com for details.

SALMONELLA TEST RESULT:

The product has been shown to be Salmonella negative using procedures recommended in the Microbiology Laboratory Guidebook, published by the USDA Food Safety and Inspection Service.

The purpose of this document is to assure that the Finished Product conforms to the above specification.

Signature: _____



Quality Control Department

Date: 3/16/2006